

**EMBEDDED DUAL BAND RFID BASED
BLOOD GLUCOSE MONITORING SYSTEM
FOR
INTERNET OF MEDICAL THINGS**

SHABINAR BINTI ABDUL HAMID

UNIVERSITI SAINS MALAYSIA

2020

**EMBEDDED DUAL BAND RFID BASED
BLOOD GLUCOSE MONITORING SYSTEM
FOR
INTERNET OF MEDICAL THINGS**

by

SHABINAR BINTI ABDUL HAMID

**Thesis submitted in fulfilment of the requirements
for the degree of
Doctor of Philosophy**

March 2020

ACKNOWLEDGEMENT

In the name of Allah, the Most Gracious and the Most Merciful

Alhamdulillah, all praises to Allah for the strengths and His blessing in completing this thesis. Special appreciation goes to my supervisor, Professor Dr. Widad Ismail, for her guidance and constant support. Her invaluable assistance with positive remarks and suggestions throughout the experimental and thesis works have contributed to the success of this research. Not forgotten, my appreciation to my co-supervisor, Associate Professor Dr. Nor Muzlifah Mahyuddin. My acknowledgement also goes to all the technicians and office staffs of School of Electrical Engineering for their cooperation.

A sincere thank you to all my friends, particularly Samihah, Solah, Rafidah and Zuraida for their kindness and moral support. My deepest gratitude to my parents, Mr. Abdul Hamid Bin Saad and Mrs. Aminah Binti Osman, as well as to all my brothers and sisters for their prayers and encouragement. I'm thankful to my husband, Shaharani Bin Husin and my children, Afiq, Haziq and Amani for their love and understanding.

I also want to thank my sponsor, the Ministry of Education and UiTM for offering me the opportunity to pursue my studies in this field. Lastly, to those who indirectly contributed in this research, your kindness means a lot to me. Thank you so much.

TABLE OF CONTENTS

ACKNOWLEDGEMENT	ii
TABLE OF CONTENTS	iii
LIST OF TABLES	viii
LIST OF FIGURES	x
LIST OF ABBREVIATIONS	xvii
ABSTRAK	xxi
ABSTRACT	xxiii
CHAPTER 1 INTRODUCTION	1
1.1 Background and Motivation	1
1.2 Problem Statements	6
1.3 Research Objectives	7
1.4 Research Scopes and Limitations.....	8
1.5 Research Contributions	9
1.6 Thesis Outline	10
CHAPTER 2 LITERATURE REVIEW	12
2.1 Introduction	12
2.2 Overview of Diabetes Mellitus.....	13
2.2.1 Patient Identification Issues in Healthcare.....	15
2.3 Internet of Things Applications and Benefits	16
2.4 Radio Frequency Identification (RFID)	19
2.4.1 Passive RFID.....	21
2.4.2 Active RFID	23
2.4.3 RFID Applications in Healthcare.....	25
2.5 Wireless Sensor Network	28
2.5.1 ZigBee (IEEE 802.15.4).....	30

2.5.2	IEEE 802.15.4 MAC Protocol	33
2.6	Related Works on the Wireless Blood Glucose Monitoring System	37
2.6.1	Knowledge and Technology Gap.....	46
2.7	Related Works on Patient Identification Mechanism.....	46
2.7.1	Knowledge and Technology Gap.....	49
2.8	Related Work on the NBE Mode in IEEE 802.15.4 MAC Protocol.....	49
2.8.1	Knowledge and Technology Gap.....	51
2.9	Design of Experiment Concept and Related Works.....	52
2.9.1	Knowledge and Technology Gap.....	56
2.10	Summary	57
CHAPTER 3 METHODOLOGY		60
3.1	Introduction	60
3.1.1	Electronic Product Code (EPC) of the RFID Tag.....	61
3.2	Overall Proposed System Design	64
3.3	Overview of the System's Description.....	67
3.3.1	Embedment of Dual Band RFID Communication Platforms	74
3.3.2	M2M Communication Realisation.....	77
3.4	DRBGM System in the IoT-WSN Platform	77
3.4.1	Microcontroller board	82
3.4.2	Embedment of Passive RFID Reader.....	84
3.4.3	Embedment of Glucometer	88
3.4.4	Power Management Design for WSN Efficiency	93
3.4.5	Proposed RF Wireless Module	95
3.4.5(a)	Operational Modes.....	98
3.4.5(b)	Application Programming Interface Packet Structure ...	99
3.4.5(c)	Acknowledgement Packet.....	101
3.5	Performance Enhancement of the NBE mode (Unslotted CSMA/CA) Design Approach	102

3.6	WSN Realisation	109
3.7	Proposed DOE Method	112
3.8	Propagation Loss	113
3.9	Summary	114
CHAPTER 4 DEVELOPMENT AND IMPLEMENTATION OF DRBGM SYSTEM.....		115
4.1	Introduction	115
4.2	Architecture Design of the Proposed DRBGM Tag.....	122
4.2.1	Physical Layer Design	122
4.2.2	Software Development.....	126
4.2.2(a)	Passive RFID Reader Communication	127
4.2.2(b)	Retrieval and Processing of Blood Glucose Information... ..	129
4.2.2(c)	Development of EN-CSMA/CA Mechanism	131
4.2.2(d)	System Communication Method of DRBGM Tag	133
4.3	Architecture Design of the Proposed Monitoring Station	134
4.3.1	Physical Layer Design of DRBGM Reader	135
4.3.2	Software Development.....	139
4.3.2(a)	Software Architecture of DRBGM Reader.....	139
4.3.2(b)	Development of the Proposed Web server.....	141
4.4	M2M Implementation in the Proposed System.....	145
4.5	Current Consumption Measurement	146
4.5.1	Evaluation and Analysis Method of Current Consumption.....	150
4.6	Transmission Range Investigation	154
4.7	Evaluating the Performance of the DRBGM system in the WSN Platform	155
4.7.1	Throughput and Latency Test Setup	156
4.8	Performance Evaluation of the Enhancement Unslotted CSMA/CA in Comparison to the Conventional Unslotted CSMA/CA	156

4.8.1	Factor Analysis for Conventional CSMA/CA Mechanism.....	156
4.8.2	Test Design for Conventional CSMA/CA Mechanism Evaluation.....	158
4.8.3	Test Design for EN-CSMA/CA Mechanism Evaluation.....	163
4.9	Performance Comparison of the Proposed DRBGM System, Conventional BGM System and Bluetooth BGM System.....	165
4.9.1	Proof-of-Concept Implementation of the DRBGM System in Comparison to the Conventional BGM System.....	165
4.9.2	Experimental setup of the Bluetooth Blood Glucose Monitoring System.....	166
4.10	Summary	167
CHAPTER 5 RESULTS AND DISCUSSION		169
5.1	Introduction	169
5.2	Performance Validation on the Proposed DRBGM System	172
5.2.1	Indoor Read Range Measurement of DRBGM system.....	172
5.2.2	Latency and Throughput Measurement and Analysis.....	176
5.2.3	Current Consumption Measurement and Analysis	180
	5.2.3(a) Evaluation and Analysis of Current Consumption	180
	5.2.3(b) Model for Estimation of the Battery Run-time	185
5.3	Analysis of Conventional CSMA/CA Anti-Collision.....	189
5.3.1	Investigation of Significant Factors Affecting Conventional CSMA/CA Performance	189
5.3.2	EN-CSMA/CA Algorithm Evaluation.....	194
5.4	Performance and Capabilities Evaluation of the Conventional, Bluetooth and the Proposed Blood Glucose Monitoring System.	197
5.4.1	Test Run at Clinic Environment.....	197
5.5	Packet Delivery Ratio Prediction Model for Integrity Proof in the Proposed Technology Deployment at the Health Facilities	204
5.6	Path Loss Prediction Model for Indoor Environment Deployment	205
5.7	Comparison between the Proposed DRBGM System and the Existing Research Works.....	209

5.8	Summary	214
CHAPTER 6 CONCLUSION AND FUTURE WORK		216
6.1	Introduction	216
6.2	Future work	217
REFERENCES		219
APPENDIX A:		
APPENDIX B: DEVELOPED GUI AND ROUTER		
APPENDIX C: CALIBRATION AND RESULT		
APPENDIX D: SOURCE CODE DRBGM TAG		
APPENDIX E: SOURCE CODE DRBGM READER		
LIST OF PUBLICATIONS		

LIST OF TABLES

	Page
Table 2.1	Blood glucose levels for diagnosing diabetes (Great Britain Association, 2017) 14
Table 2.2	Technical performance of conventional blood glucose monitoring system..... 15
Table 2.3	Healthcare applications using IoT..... 19
Table 2.4	Characteristics of RFID operating frequencies (Adhiarna & Rho, 2009)20
Table 2.5	Comparison between Gen 1 and Gen 2 (Coca & Popa, 2011).....22
Table 2.6	Active RFID versus passive RFID (Cisco, 2008)24
Table 2.7	Applications and types of RFID in a single platform26
Table 2.8	Data rates and delay demands of WBAN applications (Shu et al., 2015)29
Table 2.9	Comparison of Bluetooth, Zigbee and Wi-Fi technologies (Pothuganti & Chitneni, 2014; Kealy & Retscher, 2017)30
Table 2.10	Related works on wireless blood glucose monitoring system43
Table 2.11	Related works using DOE for analysis55
Table 3.1	Description of encoded bytes91
Table 3.2	The summary of methodologies for the proposed system 114
Table 4.1	Total current consumption for the standalone RFID tag calculated for different mode 148
Table 4.2	Total current consumption for the DRBGM tag calculated for different mode 150
Table 4.3	Summary of notations and interpretations of the standalone tag in different operations 151

Table 4.4	Summary of notation and interpretation of the DRBGM tag in different operations	152
Table 4.5	Factors and levels of DOE	157
Table 5.1	The maximum range between the tag and reader with 100% PDR .	173
Table 5.2	Comparison of the PL0 and PL4 range with 100% PDR.....	175
Table 5.3	Factors and level of DOE test	176
Table 5.4	Summary of current measurement of the standalone tag in different operations	182
Table 5.5	Summary of current measurement of the DRBGM tag in different operations	185
Table 5.6	Factors and levels of the DOE	189
Table 5.7	Factorial design table for DOE experiment	191
Table 5.8	The P-values from the ANOVA for the percentage of packet loss using the adjusted sum of squares for testing.....	192
Table 5.9	Comparison of the technical performance between the Bluetooth and proposed BGM system	197
Table 5.10	Comparison on the capabilities of the conventional, Bluetooth and proposed BGM system	199
Table 5.11	Recording time of the proposed system in comparison to conventional and Bluetooth BGM systems.....	201
Table 5.12	Patient questionnaire on the perception of the DRBGM system (n=5).....	204
Table 5.13	Transmission loss through the different types of wall	207
Table 5.14	Comparison of the proposed DRBGM system with related works. (Colour code: red for the wireless BGM system, yellow for WSN, green for anti-collision and blue for the prediction model)	211
Table 5.15	Comparison of the proposed EN-CSMA/CA to the related works ..	213
Table 5.16	Overall Specification of the DRBGM system.....	215

LIST OF FIGURES

	Page
Figure 1.1 Diabetes trends and projections by 2020 in Malaysia (Ngah et al., 2017)	1
Figure 1.2 Common place of treatment for diabetic patients (IPH, 2015)	2
Figure 1.3 Technologies associated with the IoT (Xu et al., 2014).....	3
Figure 1.4 Projected market share of dominant IoT applications by 2025 (Al-Fuqaha et al., 2015)	4
Figure 2.1 Flowchart of the literature review	12
Figure 2.2 Investments in IoT solutions by industry (Newman, 2019).....	17
Figure 2.3 Communication of IoT (Joyia et al., 2017).....	18
Figure 2.4 Passive RFID system (Tashi et al., 2011)	21
Figure 2.5 Architecture of an active RFID tag (Hamid et al., 2013).....	24
Figure 2.6 The implementation of passive and active RFID integration for an assembly line automation system (Abdullah, 2018)	27
Figure 2.7 System architecture of a typical wireless sensor node (Farooq & Kunz, 2011).....	28
Figure 2.8 ZigBee protocol stack architecture (ZigBee Alliance, 2012).....	31
Figure 2.9 Zigbee topologies	32
Figure 2.10 IEEE 802.15.4 MAC operational modes (Cunha et al., 2007).....	33
Figure 2.11 Data transmission flow chart of a BEn mode of IEEE 802.15.4 (Cunha et al., 2007).....	34
Figure 2.12 Beacon frame format (Akbar et al., 2016)	35
Figure 2.13 Data transmission flow chart of the NBE mode (unslotted CSMA/CA) of IEEE 802.15.4 (Cunha et al., 2007).....	36

Figure 2.14	Evolution of glucose sensors from invasive to non-invasive electrochemical (Lee et al., 2018).....	38
Figure 2.15	System design based on e-health sensor platform (Rasyid et al., 2016)	42
Figure 2.16	Prototype of RFID reader for reading patient’s RFID wristband (Aguilar et al., 2006)	47
Figure 2.17	The support system structure (Huang et al., 2005)	48
Figure 2.18	DOE workflow (GE Healthcare, 2014).....	52
Figure 2.19	Model of experimental design (Olasupo & Olasupo, 2017)	53
Figure 3.1	Overall research methodologies of the proposed system.....	61
Figure 3.2	Customised patient identification number via a passive RFID tag	62
Figure 3.3	Flowchart procedure for writing the EPC number to a passive tag ...	63
Figure 3.4	GUI for read/write passive RFID tag	64
Figure 3.5	Structure diagram of DRBGM system.....	65
Figure 3.6	Data recording process of conventional versus the DRBGM system at the healthcare provider	69
Figure 3.7	Overall system architecture components	70
Figure 3.8	Capability of each segment of the DRBGM system	72
Figure 3.9	Protocol integration in each component of the proposed system.....	74
Figure 3.10	Proposed embedment of the dual RFID platform into the blood glucose monitoring system.....	75
Figure 3.11	Communication between the passive RFID reader and passive RFID tag of the DRBGM system.....	76
Figure 3.12	Block diagram of the existing embedded (a) RFID tag and (b) RFID reader of PAR system (Abdullah, 2018).....	79
Figure 3.13	Block diagram of the proposed DRBGM tag.....	79
Figure 3.14	The circuit diagram of the RTC, external memory and LCD connection to the microcontroller board	80

Figure 3.15	Block diagram of the proposed DRBGM reader.....	81
Figure 3.16	Block diagram of a proposed DRBGM router	82
Figure 3.17	Custom Arduino setup for the operating clock modification of Arduino Mega	84
Figure 3.18	The circuit diagram of the passive RFID reader connected to the microcontroller board.....	85
Figure 3.19	Principle of embedded passive RFID communication.....	85
Figure 3.20	Method to issue a command string to the passive RFID reader.....	86
Figure 3.21	Customised configuration parameter of a passive reader	87
Figure 3.22	The circuit diagram of the glucometer connection to microcontroller board.....	89
Figure 3.23	Algorithm design flow chart for accessing the memory of the glucometer.....	90
Figure 3.24	Extracted encoded glucose data from EPI	91
Figure 3.25	Algorithm design flow chart of decoding of the blood glucose reading and types of meal flag	92
Figure 3.26	The sketch for date and time conversion of blood glucose data	93
Figure 3.27	Illustration of different voltage requirement of the main component.....	94
Figure 3.28	Power management solution for the proposed DRBGM tag	94
Figure 3.29	Zigbee IEEE 802.15.4 protocol stack architecture.....	96
Figure 3.30	Connection of the Zigbee transceiver to the microcontroller board for the proposed DRBGM tag and reader	97
Figure 3.31	Operational modes of Xbee Pro S2B (Digi International, 2016).....	98
Figure 3.32	Developed API transmit request frame structure for the proposed DRBGM system tag.....	100
Figure 3.33	Illustration of RF data versus ACK packet traverse path.....	101
Figure 3.34	The scenario of data collision	103

Figure 3.35	Flow chart of the conventional CSMA/CA mechanism (IEEE, 2016)	104
Figure 3.36	EN-CSMA/CA based on cross-layer interaction	106
Figure 3.37	Flow chart of the proposed EN-CSMA/CA mechanism.....	108
Figure 3.38	Configuration parameters for the Zigbee router API using XCTU software	109
Figure 3.39	Method to establish a wireless sensor network for the proposed system.....	111
Figure 3.40	Framework for developing operational guidelines of the proposed DRBGM system	112
Figure 4.1	The DRBGM system deployment architecture	115
Figure 4.2	Flowchart of the development and implementation of the DRBGM system summarized in (a) Stage 1, (b) Stage 2, (c) Stage 3, (d) Stage 4, (e) Stage 5 and (f) Stage 6	121
Figure 4.3	Block diagram of the proposed DRBGM tag architecture.....	123
Figure 4.4	Schematic diagram of the proposed DRBGM tag.....	125
Figure 4.5	PCB fabrication of DRBGM tag	126
Figure 4.6	Flowchart of passive RFID communication algorithm.....	128
Figure 4.7	Flowchart of the proposed retrieval and processing algorithm of blood glucose information.....	130
Figure 4.8	Flow chart of the proposed anti-collision scheme in the DRBGM system.....	132
Figure 4.9	Proposed algorithm flowchart of the DRBGM tag	133
Figure 4.10	Basic structure of the monitoring station	134
Figure 4.11	Block diagram of proposed RFID reader architecture	135
Figure 4.12	Schematic diagram of the proposed DRBGM reader	136
Figure 4.13	Prototype of the proposed DRBGM reader.....	137
Figure 4.14	Sample of blood glucose data tracing in a web page	138

Figure 4.15	Proposed algorithm flow chart of the DRBGM reader	140
Figure 4.16	Flowchart of the web page development of the proposed system ...	142
Figure 4.17	PhpMyAdmin of GUI for creating the DRBGM web server database	143
Figure 4.18	A web page display of a graphical chart of blood glucose distribution based on individual performance.....	144
Figure 4.19	System process flow in realising the M2M communication.....	145
Figure 4.20	Test setup for measuring current consumption	146
Figure 4.21	Theoretical voltage and current of the standalone RFID tag	147
Figure 4.22	Theoretical voltage and current of the DRBGM tag in idle mode ...	149
Figure 4.23	Testbed arrangement in communication range measurement.....	155
Figure 4.24	Experimental setup for two-hop topology.....	156
Figure 4.25	End node antenna in (a) H-field and (b) E-field polarisation.....	157
Figure 4.26	Procedures in developing the DOE experimental setup for conventional CSMA/CA evaluation	158
Figure 4.27	Chain topology with multiple hops	159
Figure 4.28	Test arrangement for anti-collision performance evaluation in a multi-floor building.....	160
Figure 4.29	Process flowchart to verify traffic being sent/delivered through designated routers.....	161
Figure 4.30	Flowchart of the algorithm for the conventional CSMA/CA mechanism evaluation	163
Figure 4.31	Software algorithm for the EN-CSMA/CA mechanism evaluation.	164
Figure 4.32	Nodes arrangement for the Proof-of-Concept test run.....	166
Figure 4.33	Conventional blood glucose monitoring in Unit Kesihatan, USM ..	166
Figure 4.34	Bluetooth blood glucose monitoring system.....	167
Figure 5.1	Overall content of the chapter results and discussions.....	171
Figure 5.2	Average RSSI versus tag-reader distance	173

Figure 5.3	Packet delivery ratio versus RSSI.....	174
Figure 5.4	RSSI versus distance for PL4 and PL0 power transmission.....	175
Figure 5.5	The p-value from the analysis of variance using the adjusted sum of squares for throughput performance.....	177
Figure 5.6	Main effects plot for throughput performance.....	178
Figure 5.7	(a) Latency and (b) throughput performance of the embedded DRBGM system versus the standalone system.....	179
Figure 5.8	Measured current during transmission, idle and reception of the standalone tag.....	181
Figure 5.9	Measured current at (a) start up, tag scanning, ID streaming and retrieve glucose data (b) transmit mode and receive mode (c) idle and sleep mode of DRBGM tag.....	184
Figure 5.10	Expected battery run-time of the DRBGM tag versus the frequency of data transmission for three different number of retransmissions of glucose information.....	186
Figure 5.11	Mean of drain current versus interval period for a different number of retransmissions of glucose data.....	188
Figure 5.12	TrackDIA displays the data collected during conventional CSMA/CA evaluation.....	190
Figure 5.13	Main effect plot for the packet loss rate.....	193
Figure 5.14	Interaction plot for the packet loss rate.....	194
Figure 5.15	PDR of an EN-CSMA/CA in comparison to the conventional CSMA/CA algorithm.....	195
Figure 5.16	Comparative graph of latency measurement.....	196
Figure 5.17	Current consumption versus time during a Bluetooth connection...	198
Figure 5.18	Conventional (a) blood glucose monitoring system and (b) manual recording at Unit Kesihatan, USM.....	200
Figure 5.19	Blood glucose test conducted on (a) real patient at Unit Kesihatan and (b) sample of data viewed at the web portal.....	200

Figure 5.20	Flowchart of the link procedure between the Bluetooth BGM meter, CHIEF mobile application and the CHIEF web portal.	202
Figure 5.21	Blood glucose result from (a) Bluetooth blood glucose monitoring system and updated to (b) the mobile app.....	203
Figure 5.22	Comparison of PDR prediction based on the selection factors using a modelling equation	205
Figure 5.23	Linear curve fit for (a) LoS and (b) NLoS receive signal strength ..	206
Figure 5.24	Comparison chart of the empirical and experimental PLO.....	208

LIST OF ABBREVIATIONS

ACK	acknowledgement
ADAPT	ADaptive Access Parameters Tuning
ANOVA	Analysis of variance
AODV	Ad hoc On-demand Distance Vector
AP	antenna polarization
API	Application Programming Interface
APL	Application
APS	Application Support
ASK	Amplitude-shift keying
AT	Transparent
BE	Backoff Exponent
BE _n	Beacon-enabled
BI	Beacon Interval
BG	Blood glucose
BGM	Blood glucose monitoring
BGMD	Blood glucose monitoring device
BGMS	Blood glucose monitoring system
BO	Beacon Order
CAP	contention access period
CCA	Clear Channel Assessment
CSMA/CA	Carrier Sense Multiple Access with Collision Avoidance
CV	Coefficient of variation
DM	Diabetes mellitus
DOE	Design of experiment
DR	Delivery ratio
DRBGM	Dual band RFID Based Blood Glucose Monitoring
DUT	Device under test
ECC	Event Chains Computation
EEPROM	Electrically Erasable Programmable Read-Only Memory
ECG	electrocardiogram

EN-CSMA/CA	Enhancement Unslotted CSMA/CA
EPC	Electronic Product Code
FDA	Food and Drug Administration
FFD	Full-Function Devices
GND	Ground
GPIO	General-Purpose Input/Output
GSM	Global System for Mobile communications
GUI	Graphical User Interface
HTTP	Hyper Text Transfer Protocol
ID	Identification
IEEE	Institute of Electrical and Electronics Engineers
I-MAC	Hybrid MAC protocol using interrupt mechanism
IoT	Internet of Things
IoMT	Internet of Medical Things
IPH	Institute of Public Health
I2C	Inter-Integrated Circuit
ISO	International Organization for Standardization
ISM	Industrial, Scientific, and Medical
ISP	Internet Service Provider
IVGTT	Intravenous Glucose Tolerance Test
LCD	Liquid Crystal Display
LIFO	Last-In-First-Out
LoS	Line of Sight
M2M	Machine-to-machine
MAC	Medium Access Control
MAUDE	Manufacturer and User Facility Device Experience
MBDOE	Model-based design of experiment
MCMC	Malaysian Communications and Multimedia Commission
MOH	Ministry of Health
NB	Number of Backoffs
NBE	on-beacon-enabled
NHMS	National Health Morbidity Survey
NoEN	Number of end nodes
NoH	Number of hops

NLoS	Non Line of Sight
NHSLA	National Health Service Litigation Authority
NWK	Network
P2P	Point-to-point
PAN	Personal Area Network
PAR	Passive and Active RFID
PC	Personal Computer
PCB	Printed Circuit Board
PCD	Portable communication device
PD	Power Divider
PDA	Personal Digital Assistant
PDR	packet delivery ratio
PHP	Hypertext Preprocessor
PHY	Physical layer
PID	Patient identification
PLO	Path loss
PL	Power level
PLR	Packet Loss Rate
PoC	Proof of Concept
PWM	Pulse width modulation
QoS	Quality of Service
REQ	Request
RF	Radio Frequency
RFD	Reduced-Function Devices
RFID	Radio Frequency Identification
RSSI	Received Signal Strength Indicator
RTC	Real-time clock
RTF	Reader-Talk-Fist
RX	Data Receive
SA	Spectrum Analyser
SDEV	Standard deviation
SD	Superframe Duration
SIM	Subscriber Identification Module
SG	Signal generator

SN	Number of Sleep Periods
SO	Superframe Order
SP	Sleep Period
TCP	Transmission Control Protocol
THIS	Total Hospital Information System
TTF	Tag-Talk-First
TUT	tag under test
TX	Data Transmit
UART	Universal Asynchronous Receiver/Transmitter
UCSMA	Unslotted CSMA/CA probabilistic polling
UDP	User Datagram Protocol
UHF	Ultra High Frequency
UiTMPP	UiTM Pulau Pinang
USB	Universal Serial Bus
USSD	Unstructured Supplementary Service Data
USP	User Services Platform
VB	Visual Basic
WBAN	Wireless Body Area Network
WHO	World Health Organization
Wi-Fi	Wireless fidelity
WLAN	Wireless Local Area Network
WSN	Wireless Sensor Network
ZC	ZigBee Coordinator
ZDO	Zigbee Device Object
ZED	Zigbee End Device
ZR	Zigbee Router

SISTEM PEMANTAUAN GLUKOSA DARAH BERASASKAN RFID DWI JALUR TERBENAM UNTUK INTERNET KEBENDAAN PERUBATAN

ABSTRAK

Maklumat kesihatan yang direkodkan secara manual boleh menyebabkan kesilapan seperti pengenalan pesakit dengan tidak tepat dan salah memadamkan data pesakit yang boleh menjejaskan keselamatan pesakit dengan serius. Untuk mengurangkan risiko kesilapan kepada pesakit diabetes, satu reka bentuk sistem pemantauan glukosa darah tanpa wayar dengan pembenaman RFID dwi jalur untuk Internet Kebendaan Perubatan telah dibangunkan. Dengan menggunakan kaedah ini, RFID pasif membolehkan komunikasi jarak dekat untuk membaca nombor pengenalan pesakit dan RFID aktif memanjangkan komunikasi jarak jauh untuk merekod dan memantau data glukosa darah wayar melalui rangkaian berbilang hop. Kerja yang dibentangkan dalam tesis ini menyumbang terutamanya kepada sistem terbenam dan aplikasinya dalam penjagaan kesihatan untuk mengurangkan beban merekod, mengesan dan memantau data pesakit dengan membenamkan sensor glukosa darah, RFID pasif, RFID aktif, WSN, M2M dan IoMT ke dalam satu platform. Konsep reka bentuk baru diwujudkan untuk mekanisme pengecaman pesakit, di mana mekanisme itu dibenamkan ke dalam peranti sumber untuk meningkatkan keupayaan sistem menetapkan nombor pengenalan secara automatik kepada setiap pengukuran glukosa darah (mmol/L) semasa pemantauan berbilang pesakit. Selain itu, keputusan dari eksperimen yang dijalankan menunjukkan bahawa sistem yang dibangunkan menghasilkan prestasi keseluruhan yang lebih baik berbanding dengan system BGM Bluetooth dan sistem BGM konvensional dari segi masa rakaman terpendek dan keupayaan penghantaran semula data. Dalam analisis

kebolehpercayaan dengan menggunakan kaedah statistik ANOVA dan DOE, mengesahkan bahawa bilangan hop dan bilangan nod akhir memberikan kesan ketara kepada prestasi PDR CSMA/CA konvensional. Kedua-dua parameter ini kemudiannya diambil kira dalam persediaan eksperimen untuk menilai prestasi algoritma CSMA/CA yang dipertingkatkan (EN-CSMA/CA), yang menggunakan mekanisme interupsi luaran dan pendekatan berasaskan lapisan silang. PDR meningkat daripada 94% (CSMA/CA konvensional) kepada 99.33% (EN-CSMA/CA), iaitu peningkatan 5.33%. Model PDR menganggarkan bahawa untuk senario terbaik dan paling teruk, peratusan PDR adalah 100.0% dan 51.67%, masing-masing. Untuk mengoptimumkan penyusunan penghala bagi pelaksanaan sebenar sistem DRBGM di kemudahan kesihatan, model kehilangan laluan yang dibangunkan menganggarkan bahawa penghala harus diletakkan pada jarak 30 m antara satu sama lain, yang disepakati dengan hasil ujian yang menunjukkan kedudukan penghala harus ≤ 40 m untuk mencapai prestasi terbaik PDR. PDR meningkat daripada 94% (CSMA/CA konvensional) kepada 99.33% (EN-CSMA/CA), peningkatan 5.33%. Model PDR menganggarkan bahawa untuk senario terbaik dan paling teruk, peratusan PDR adalah 100.0% dan 51.67%, masing-masing. Untuk mengoptimumkan penyusunan penghala bagi pelaksanaan sebenar sistem yang dibangunkan di kemudahan kesihatan, model kehilangan laluan yang dibangunkan menganggarkan bahawa penghala harus diletakkan pada jarak 30 m antara satu sama lain, yang disepakati dengan hasil ujian yang menunjukkan kedudukan penghala harus ≤ 40 m untuk mencapai prestasi terbaik PDR.

EMBEDDED DUAL BAND RFID BASED BLOOD GLUCOSE MONITORING SYSTEM FOR INTERNET OF MEDICAL THINGS

ABSTRACT

Manually recorded health information could lead to errors such as inaccurate patient identification and mismatch patient data that could seriously affect patient safety. In order to reduce the risks of error for patients with diabetes, a new design of wireless blood glucose monitoring system with the embedment of dual band RFID for Internet of Medical Things is being developed. Using this method, passive RFID allows short-range communication to read automatically the patient identification number and active RFID extends long-range communication for recording and monitoring blood glucose data through multi-hop WSN. The work presented in this thesis contributes mainly to the embedded system and its application in healthcare to reduce the burden of recording, tracing and monitoring the patient's data by embedding blood glucose sensor, passive RFID, active RFID, WSN, M2M and IoMT into a single platform. A new design concept is established for the patient identification mechanism, where the mechanism is embedded in the source device to enhance the ability of the system to automatically assign the identification number to each blood glucose measurement (mmol/L) during multiple patients monitoring. Additionally, the results from the experiments conducted showed that the developed system produced better overall performance compared to the Bluetooth BGM and conventional BGM system in terms of the shortest recording time and the ability to retransmit data. In the reliability analysis using ANOVA and DOE statistical methods, the result validates that the number of hop and number of end node significantly affects the PDR performance of conventional CSMA/CA. These two

parameters are then taken into account in experimental setup for performance evaluation of the enhanced CSMA/CA (EN-CSMA/CA) algorithm that uses an external interrupt mechanism and a cross layer approach. The PDR increased from 94% (conventional CSMA/CA) to 99.33% (EN-CSMA/CA), an improvement of 5.33%. The PDR model estimates that for the best and worst scenario, the percentage of PDR is 100.0% and 51.67%, respectively. To optimize the arrangement of routers for real implementation of the developed system in health facilities, the developed path loss model estimates that the router should be positioned at a distance of 30 m from each other, which agrees with the test results which indicate that the router should be positioned ≤ 40 m in order to achieve the best PDR performance.

CHAPTER 1

INTRODUCTION

1.1 Background and Motivation

Diabetes mellitus (DM) is a significant public health problem that is recognised globally. The World Health Organisation (WHO) estimates that by the year 2030, Malaysia will have around 2.48 million people with DM (IPH, 2011). Figure 1.1 illustrates the trends and projections of diabetes by the year 2020, showing that the prevalence of diabetes even in 2011, based on the National Health Morbidity Survey (NHMS), was higher compared to the projected prevalence of diabetes in NHMS 2006 (Ngah et al., 2017).




Figure 1.1 Diabetes trends and projections by 2020 in Malaysia (Ngah et al., 2017)

NHMS 2015 revealed that most people with diabetes sought medical treatment at Ministry of Health (MOH) health clinics (59.3%), followed by MOH hospitals (20.0%), private clinics (15.1%) and private hospitals (3.6%) as illustrated

in Figure 1.2 (IPH, 2015). NHMS 2015 also estimated that the high population growth in Malaysia, if continued, will undoubtedly overload the nation's healthcare systems. Currently, the ratio of doctors to the population is 1:633 and the ratio of nurses to the population is 1:333 (Mei Kei, 2017).




Figure 1.2 Common place of treatment for diabetic patients (IPH, 2015)

The MOH has implemented a Total Hospital Information System (THIS) in order to improve the scheduling and day-to-management of surgical operations and services within the healthcare system. In utilising this system, diabetes health records should be recorded directly into the system electronically, and information and equipment tracked through using barcode identification tags and patient ID numbers. At present, nurses record blood glucose data on a piece of paper, and following the conclusion of their ward rounds, will enter the data into the hospital system at the nurses' station. In contrast, the proposed process as outlined in this study will provide a significant benefit concerning the reliability of blood glucose data compared to manually recorded information which may not be accurately entered and recorded into the system. For instance, data may become mixed up with other data during the

data entry and compiling process in the hospital system. Therefore, to improve the reliability and integrity of blood glucose data, the use of a wireless blood glucose monitoring (BGM) system will assist to automatically store data into THIS, thus minimising the occurrence of errors due to human intervention.

Also important is the accurate identification of patients in the effective management of diabetes. From patient admission and throughout the range of care provided by the hospital, matching the correct patient name and personal details to the correct medication, specimen, test, and procedure are crucial to improving patient safety. In the conventional manual process to identify patients, mistakes can invariably occur which may seriously compromise the patient's safety and health condition, and preventing them from receiving proper treatment.

The emergence of smart devices, Radio Frequency Identification (RFID), Wireless Sensor Network (WSN) and other communication technologies as shown in Figure 1.3 has led to the exponential growth of online applications associated with the Internet of Things (IoT).




Figure 1.3 Technologies associated with the IoT (Xu et al., 2014)

In fact, it has been projected that by the year 2025, healthcare applications will dominate the IoT as shown in Figure 1.4. In the modern health care environment, the application of IoT technologies has brought with it, the convenience for both medical professionals and patients, given these applications can be applied to various medical areas and procedures including real-time monitoring, patient information management, and healthcare management. Indeed, electronically connected medical devices will help to free up much-needed resources in clinics, alleviate and reduce stress and costs for those undergoing treatment, and will ultimately improve the delivery of services in hospitals and other health care facilities.




Figure 1.4 Projected market share of dominant IoT applications by 2025 (Al-Fuqaha et al., 2015)

For indoor healthcare monitoring systems, several alternate technologies have been proposed including the use of Wi-Fi, RFID and Bluetooth. Mainetti et al. (2014) carried out a comparison among these technologies and concluded that RFID and Bluetooth technologies were the optimal solutions to reduce costs and providing less complexity in terms of hardware development and implementation. Blood glucose testing in hospitals and clinics involves a multitude of patients, covering a large area in monitoring patients simultaneously. The main concern in using a Bluetooth glucometer in a health facility for blood glucose monitoring is due to the limitation in transmitting data over long distances, inability to provide unique identification for multiple patients and compatibility with smartphone devices. Additionally, the pairing process of the Bluetooth glucometer requires a lengthy setup process, whereas RFID can provide an effective solution in overcoming the limitations of using Bluetooth technology.

Embedding RFID and WSN technology in forming IoT components can contribute to remote and automated blood glucose monitoring in a health facility environment. Furthermore, the deployment of a diabetes monitoring system utilising the IoT platform will enable the system to be deployed anywhere. For remote monitoring, doctors can access data via an online diabetes management system for tracking the performance of patients' blood glucose levels which can assist doctors in diagnosing and providing treatment to patients. Therefore, there is a need to improve the BGM system by utilising RFID in a WSN platform and embedding the system with a mechanism to identify patients for crowd testing and regular usage. To the best of the researcher's knowledge, such a design of a wireless BGM system is the first of its kind.

1.2 Problem Statements

In crowd blood glucose monitoring, the existing wireless BGM system does not have the ability to automatically attach a patient identification number to the blood glucose reading for each patient during the remote recording of data into an electronic health record system. To overcome the limitations, the proposed system requires wireless technology capable of managing wireless communication at both short and long ranges. The short-range is for collecting the identification number of the patient, while the long-range is for remote data recording. Nonetheless, the wireless technology cited in the literature for the existing blood glucose monitoring system operates in a single band, restricting system interoperability and functionality to manage short- and long-range application. Besides, the existing wireless BGM design is focused on self-monitoring instead of monitoring multiple patients in which a single device can be shared across multiple patients. For example, the Bluetooth BGM system invented by Simpson et al. (2010) and Ow-Wing (2015), the identification number of the patient is registered in the mobile application at the receiving device. Such design limits the BGM system in detecting the identification number of different patients, as it requires the removal of the previous patient's data by uninstalling the mobile application and reinstalling it. Moreover, data collection and data analyses are dependent on human's intervention to input the data.

The main concerns regarding communication schemes for a BGM system in the WSN platform is regarding the ability to use the system remotely in real-time and to preserve the integrity of the monitoring data. Zigbee provides solution to data collision with the use of CSMA/CA mechanism. Due to its technical uniformity, highly scalability, the efficiency of CSMA/CA mechanism of ZigBee protocol will be affected in specific environments. In fact, collision rate and back off period are

still relatively high and long in star topology deployment even though the Zigbee network occupied only with fewer nodes (Singh, 2012). The development of the beacon-enabled mode (slotted CSMA/CA) protocol developed by Shu et al. (2015) and Di Francesco et al. (2011) is not feasible in accommodating the need for real-time data processing in a multi-hop network as the nodes need to synchronise with the coordinator and listen to periodic beacons in order to the transmit data. In contrast, using an unslotted CSMA/CA mechanism, the nodes have no restriction in transmitting data at any time due to no alignment to a slot boundary. However, if all nodes attempt to transmit data simultaneously, data will collide which will jeopardise the integrity of collected data and affect the analysis of data for diagnosis and treatment of diabetic patients. The majority of work on the development of the unslotted CSMA/CA protocol has only considered simulation analysis, for example, the work developed by De Guglielmo et al. (2016) and Ur Rehman et al. (2014). These works also do not provide analysis on the multi-hop network given this information is vital for prediction efficiency in implementing the system in a large area of a health facility.

1.3 Research Objectives

The objectives of the research are as follows.

- i. To improve the design and functionalities of the wireless BGM system by embedding dual band RFID (passive and active RFID) into a single WSN platform by focusing on patient identification number assignment to the blood glucose reading for each patient during remote data recording into an electronic system in order for the system to be used in crowd monitoring with regular use in a large area of a health facility.

- ii. To propose a new approach to enhance conventional unslotted CSMA/CA for preserving data integrity, to evaluate the performance of the proposed protocol and improve the reliability of real-time wireless blood glucose monitoring data.
- iii. To develop and construct a prototype of the proposed real-time wireless BGM system as a proof of concept for the implementation of the Internet of Medical Things.

1.4 Research Scopes and Limitations

The research work in this thesis focuses on two elements. First by defining the problem precisely and proposing solutions, and secondly, by producing meaningful results in order to achieve the objectives of this research.

The first element incorporates the process of designing and developing embedded dual-band RFID (passive and active) tags and the active RFID reader starting from designing the circuit diagrams, selection of the glucometer, passive reader, microcontroller, active components, design circuit schematic, the Printed Circuit Board (PCB) layout design and fabrication as well as installation of the components.

The second element encompasses software development, which focuses on various functions for the microcontroller to communicate with the glucometer, passive reader, Zigbee module, external memory, in real-time as well as developing the GUI to interface with the RFID reader. The tag is designed not only to communicate with the glucometer and passive reader but also to achieve high data integrity of glucose information by embedding the enhancement anti-collision algorithm into the program coding. The raw data sent by the system tags contain

information such as the IEEE address, patient ID, glucose reading, time stamp, meal flags and patient comment. The developed system supports M2M communication that enables the active RFID reader to exchange real-time information with the embedded active RFID tags in a two-way communication channel to update the status of delivery and validate the data with minimal human intervention through utilising the IoT environment.

All experiments are conducted in a real-world environment by assuming that the same environment is applied during blood screening of diabetic patients to ensure the research objectives are achieved. However, this work does not focus on the accuracy of the blood reading as the glucometer that is used in this research is a commercial device. Instead, the focus is towards system design, development, improvement, implementation, embedding technology and testing in order to achieve the objectives of this research from various perspectives.

1.5 Research Contributions

The contributions of this research are summarised as follows.

- i. New design architecture comprises of embedded blood glucose sensor, passive RFID, active RFID, WSN, M2M and IoMT in a single platform for real-time wireless blood glucose monitoring for healthcare facilities. This new embedded system enables short-range communication to read patient identification number and extend long-range communication via multi-hop network for remote blood glucose data recording into an electronic system. A new design concept is established for a patient identification mechanism, where the mechanism is embedded into a source device. This enhances the

capability of the proposed system to be used for multiple patients, and each blood glucose measurement is automatically assigned to their identification number during blood glucose testing. These features allow easy tracking, monitoring and analysis.

- ii. Formulating new approaches and algorithm for performance enhancement unslotted CSMA/CA (EN-CSMA/CA) for preserving the integrity and improving the reliability of real-time monitoring data of the developed BGM system in large area of health facility. This is achieved by implementing a DOE approach and statistical analysis to verify the robustness of the system and performance and also considering multi-hop network factor analysis.
- iii. A new prototype system is developed and implemented as proof of concept in performing M2M communication for remote blood glucose monitoring. This embedded system is in the form of a portable wireless BGM system which is relatively easy to be set up in a health facility thereby allows more extensive wireless coverage in a multi-hop network.
- iv. Formulating a packet delivery ratio model for predicting the effects from the distance and interaction factors of a number of DRBGM tags * number of hops, for reliable implementation in large area of health facility.

1.6 Thesis Outline

The structure of this thesis is comprised of six chapters as described below.

Chapter Two reviews the various methodologies and problems with the current wireless BGM system based on existing literature.

Chapter Three describes the specifications and operation of the key components used in the research work in order to develop an active RFID system

embedded with a glucose meter and passive reader functionality with the aid of a WSN communication platform for indoor monitoring. This chapter also describes the hardware and software design, development and configuration of a complete embedded RFID system.

In Chapter Four, the methodology used to setup the hardware, software and network design and development including the embedded mechanism between devices and WSN in the IoT is described. The schematics of the embedded RFID tag, reader and router are presented along with the communication algorithm implemented in the RFID tag and reader, respectively. All testing and results concerning the proposed embedded RFID system prototype are compared with the standalone RFID system and an RFID system embedded with sensors to evaluate and verify the capabilities of the proposed system.

Chapter Five describes the findings and discusses the process of testing the wireless BGM system as a Proof of Concept (POC) in the selected real-world indoor environment. This chapter also details the actual contributions and advantages of the entire package of the proposed embedded RFID system by summarising the findings in the form of design specifications for the RFID tag and reader together with a comparison between the proposed embedded RFID system with prior approaches found in the literature.

Chapter Six concludes the research by addressing the objectives of the research along with recommendations for future improvements.

CHAPTER 2

LITERATURE REVIEW

2.1 Introduction

The chapter begins with an overview of diabetes mellitus and continues by highlighting the patient identification issues in healthcare. Discussion on the important aspects of Internet of Things (IoT) applications, RFID, WSN and Zigbee are next presented along with a comparison with other wireless technologies. This chapter next examines the existing body of literature on the design of a wireless blood glucose monitoring (BGM) system and related works on patient identification technologies. Related to the event-driven application as implemented in the proposed system, the chapter examines the issue of IEEE 802.15.4 MAC by investigating the related works in addressing the issue. Design of experiment concept is discussed in the last section of this chapter. The flowchart of the literature review is summarised and illustrated in Figure 2.1.




Figure 2.1 Flowchart of the literature review

2.2 Overview of Diabetes Mellitus

The Egyptians first mentioned diabetes mellitus around 1500 BCE. Diabetes mellitus (DM) is an incurable disease resulting from an insufficient amount of insulin in the human body, causing elevated blood glucose levels, known as hyperglycaemia, or reduced glucose concentrations, known as hypoglycaemia (Sabokdast et al., 2015). Insulin is a hormone made by the pancreas in the body that regulates the amount of glucose in the blood.

Diabetes is classified into type 1 diabetes, type 2 diabetes, and gestational diabetes. Type 1 diabetes occurs when no or very little insulin is released into the body due to the immune system mistakenly attacking and killing the beta cells of the pancreas. As a result, sugar builds up in the blood instead of being used as energy. Type 1 diabetes requires treatment with insulin. Whereas, type 2 diabetes occurs when the body cannot properly use the insulin that is released or does not make sufficient insulin. Similar to type 1, sugar builds up in the blood instead of being used as energy. Type 1 and type 2 diabetes have been associated with many medical complications including cardiovascular disease, neuropathy, nephropathy, retinopathy, foot damage, skin problems, hearing impairment and chronic depression. The third type of diabetes, gestational diabetes, is a temporary condition that occurs during pregnancy. Here, uncontrolled blood sugar levels in pregnancy can also cause the baby to grow too large, having a higher risk of developing obesity as well as causing death either before or shortly after birth. Additionally, the mother may also experience preeclampsia and subsequent gestational diabetes (Berger et al., 2016).

Complications from diabetes have resulted in greater expenditure and reduced productivity, becoming a socio-economic concern (Huse et al, 1989; O'Brien et al., 1998). Moreover, monitoring glucose levels in the blood has been proven to prolong

life expectancy by enabling people with diabetes to manage episodes of hypo- or hyperglycaemia, therefore providing better control over their condition and by preventing some of the debilitating side effects (Makaram et al., 2014; Ferrante do Amaral & Wolf, 2008).

The key part of diabetes management is in understanding blood glucose level ranges, which are checked using a blood glucose monitoring system. The blood sugar levels in diagnosing diabetes are shown in Table 2.1.

Table 2.1 Blood glucose levels for diagnosing diabetes (Great Britain Association, 2017)

Plasma glucose test	Normal (mmol/L)	Prediabetes (mmol/L)	Diabetes (mmol/L)
Random	Below 11.1	N/A	11.1 or more
Fasting	Below 5.5	5.5 to 6.9	7.0 or more
2 hour post-prandial	Below 7.8	7.8 to 11.0	11.0 or more

According to the results from the 2011 National Health & Morbidity Survey (NHMS), an estimated 1.1 million patients with diabetes received treatment at public health care facilities in Malaysia, with 70% of patients attending primary care clinics. Whereas, the remaining patients received treatment and follow-up at public hospitals (IPH, 2012). Notably, medical officers and physicians mostly deliver public hospital-based diabetes care within general medicine outpatient clinics in the hospital (Hussein et al., 2015).

On the other hand, glycaemic control for hospitalised patients with diabetes requires accurate near-patient glucose monitoring systems. During the last decade, point-of-care blood glucose monitoring devices have become the mainstay of near-patient glucose monitoring in hospitals globally (Rajendran & Rayman, 2014), given that the blood glucose measurement result can be obtained in minutes compared to what could take hours employing laboratory tests. The Ministry of Health in

Malaysia has enforced the use of Total Hospital Information Systems (THIS) to record all health-related data, including blood glucose data into the system. However, the data entry of THIS is a manual process, which is susceptible to human errors. The technical performance of conventional blood glucose monitoring (BGM) system is tabulated in Table 2.2. Therefore enhancing the functionality and capability of the BGM system could positively affect the level of work involved and enhance patient care across the public health care facilities.

Table 2.2 Technical performance of conventional blood glucose monitoring system

System	Conventional BGM System
Method to record data to database system	Manual data entry
User identification mechanism	Visual check the Patient ID Card
Type of blood glucose data collection	Non-real-time
Number of patient per system setup	One patient
Diabetes Management Software	NA

2.2.1 Patient Identification Issues in Healthcare

A blood glucose reading without knowing the identification of a patient is meaningless. Patient identification refers to the process of assigning a unique identification ID to a patient. Patient identification is one of the most important elements of patient safety by identifying each patient precisely and correctly matching their data to facilitate the best possible decisions about treatment and diagnosis (Morris, 2014). The Joint Commission for Accreditation of Health Care Organizations (2016) recommends the use of at least two identifiers (e.g., the patient's name and their date of birth) to verify a patient's identity. In the health facilities in Malaysia, the inpatients are identified based on the presence of the ID

band that contains information on the patient's full name, medical record number and date of birth whereas, outpatients are identified through their National ID and are confirmed with details contained in their medical record.

However, there is significant evidence that errors in the identification of patients do occur in health services. The misidentification of patients has been disclosed as the root cause for the majority of medical errors including specimen misidentification (Upreti et al., 2013), adverse drug events (Fyhr & Akselsson, 2012) and incorrect blood transfusion (Cottrell et al., 2013), which can cause the loss of a patient's life. In the root cause analyses study, Dunn and Moga (2010) found that 9.48% out of 182 patient ID errors in the medical results from the post-analytic phase were recorded into the wrong patient medical record. While, the study conducted at the Endocrinology-Diabetology-Nutrition Department at Montpellier Hospital, France revealed that patients with diabetes had more serious medication errors at the hospital compared to non-diabetic patients (Breuker et al., 2017). Medication errors can be lethal, with the estimated death of more than 7,000 patients each year from preventable errors (Walsh, 2015). In Malaysia, the figures in 2010/2011 NHSLA Annual Report and Accounts reveals that the number of clinical negligence claims reported in 2010/11 was 8655, which represents a 30% increase over 2009/10 (6652) (Hambali & Khodapanahandeh, 2014).

2.3 Internet of Things Applications and Benefits

The IoT comprises a network of electronic physical devices that are interconnected, exchange data and interacts with each other beyond machine-to-machine (M2M) communication. Multiple technologies are involved in the IoT, with one solution in managing all communications. This ingenious evolution has attracted

the interest and attention of many industries to be part of the IoT such as information, transportation, agricultural, healthcare and manufacturing industries as shown in Figure 2.2. In balancing the demand and supply in healthcare services, the growth of the IoT continues to grow and proliferate. The capability of medical devices and application systems to interact and share data through online computer networks enable fast, accurate, seamless and transparent services in the healthcare system.



Figure 2.2 Investments in IoT solutions by industry (Newman, 2019)

In other spectrums, the industrial revolution aimed at creating intelligent factories by applying advanced technologies including the IoT and cloud computing is known as Industry 4.0 (Lee et al., 2015). In the context of the healthcare system, the term Industry 4.0 is used to represent an in-depth change in the way healthcare services are provided and how information, technology and processes integrate forming a lean healthcare system. Lean healthcare is a set of operating philosophies and methods that help to create maximum value for patients by reducing waste and

waiting time (Lawal et al., 2014). Hence, integrating Industry 4.0 in the healthcare system is important given the benefits of improving access to care, reducing healthcare costs and most importantly increasing patients' quality of life. This transformation can be realised through utilising and taking advantage of the IoT.

Innovative technologies like ZigBee, WSN, RFID, smart sensors, distributed databases and energy saving technologies, associated with IoT, have huge potential in remote healthcare monitoring systems and medical services to improve patient safety, healthcare services and also to reshape the future of the healthcare industry. The interaction of the elements in IoT communication is illustrated in Figure 2.3.




Figure 2.3 Communication of IoT (Joyia et al., 2017)

Given the benefit afforded by IoT infrastructure, many researchers have shown significant interest in adopting IoT in their applications. Table 2.3 shows some of the related work in healthcare applications that have adopted the IoT concept.

Table 2.3 Healthcare applications using IoT

Authors	Application	Benefit
Muhammad et al. (2017)	Voice pathology monitoring	High accuracy detection and easy to use
Chandel et al. (2016)	Continuous monitoring for step count, stride length, fall and calorie	Trigger an automatic alarm for the need of an urgent attention to the elderly subject, real-time feedback on step count and distance , accurate calories estimate
Puri et al. (2016)	Cardiac Arrhythmia Management	enable an inexpensive auto-triggered arrhythmia cardiac management solution catering the need of in-house, round-the-clock cardiac health monitoring

2.4 Radio Frequency Identification (RFID)

The concept of RFID was initially discovered during World War II (Landt ,2005) which has since evolved. RFID provides a feasible solution to all tracking applications using an identification mechanism via radio waves. The frequency ranges of RFID based on the international standards range from 135 kHz to 2.4 GHz. The characteristics of RFID operating frequencies are described in Table 2.4. RFID can be categorized into passive and active which are embedded for short and long range communication, respectively, in the proposed system. Details of these RFIDs are further discussed in the next section.

Table 2.4 Characteristics of RFID operating frequencies (Adhiarna & Rho, 2009)

Frequencies	Low Frequency (LF)	High Frequency (HF)	Ultra High Frequency (UHF)	Microwave (MW)
	125,134-135 kHz	13.56 MHz	902-908 MHz (US), 865 – 868 MHz (Europe)	2.4 GHz
Characteristics	1.Short-range (<1m) 2.Low data transfer rate (<1kbps) 3.Penetrate water but not metal	1.Higher range (<1.5m) 2.Reasonable data rate = 25 kbps (similar to GSM phone) 3.Penetrate water but not metal	1.Long-range (3-6m Europe, 6-8m US for passive tag) and >100m for active tag) 2.High data transfer rate (=30 kbps) 3.Cannot penetrate water and metal	1.Long-range (>10m and 100m for active tag) 2.High data transfer rate (>100 kbps) 3.Cannot penetrate water or metal
Operating principles	Inductive coupling		Electromagnetic coupling (backscatter)	
Energy supply	Typically passive		Typically passive and active	
Tag size	Large	Medium	Small	Small
Typical RFID Applications	Animal tracking, Access controls, Car keys	Smart cards, Access control, Public transport, Item level identification, Apparel application	Supply chain pallet & box level identification, Electronic toll collection	Finished products tracking, Electronic toll collection
ISO Standards	14223 11784/5 18000-2	15693 14443 18000-3	10374 18000-6	10374 18000-4 18000-5
EPC Standards	-	Class 1 (Gen1) HF	Class 0 (Gen1) Class 1 (Gen1) UHF Class 2 (Gen2)	-

2.4.1 Passive RFID

A basic configuration of a passive RFID system consists of a reader, passive tag, and application computer (Tashi et al., 2011) as illustrated in Figure 2.4. The type of RFID system depends on the type of utilised tag. A passive tag does not have a battery, and vice versa for an active tag. The passive tag obtains all the energy from the generated RF field and then backscatters the signal to the RFID reader (Chawla & Ha, 2007). The backscatter signal contains data in the form of an identification string, which is pre-stored in the internal memory of the passive tag.




Figure 2.4 Passive RFID system (Tashi et al., 2011)

In the majority of cases, the read range of the passive RFID system is limited given the low sensitivity of the passive tag which requires a strong RF signal from the reader (Koelle et al., 1975). Conventional passive RFID tags are limited in their capability; however, the tags have an indefinite operational life and are sufficiently tiny enough to fit into a practical adhesive label (Want, 2006).

The passive RFID systems can be operated in low frequency (LF), high frequency (HF), and ultra-high frequency (UHF) bands. However, there are currently no commercially available passive RFID ICs for 2.4 GHz (Nikitin, 2019). Numerous communication protocols exist for RFID systems as shown in Table 2.5. The

standards and guidelines for the interactions between the passive tags and the readers are based on the Electronic Product Code (EPC) global specifications (GS1, 2015) and International Organization for Standardization (ISO) 18000-6 (ISO, 2013), which are adopted by industry in the Ultra High Frequency (UHF) passive RFID field. The first generation of the RFID was named as Generation 1 (Gen 1) and the second generation (Gen 2) of the RFID UHF tags were then developed to establish a standard for RFID tags, operating frequency band and offer a longer operational range (Coca & Popa, 2011). Table 2.5 displays details of the improvements in Gen 2 in comparison to the Gen 1 protocol.

Table 2.5 Comparison between Gen 1 and Gen 2 (Coca & Popa, 2011)

Description	Gen 1	Gen 2
Acceptance level	Not a global standard	ISO-UHF 18000-6 standard
Arbitration	Deterministic binary tree for Class 0 and deterministic slotted for Class 1	Probabilistic slotted
Anti-collision/tag-sorting algorithm	Binary tree algorithm with persistent state/wake states	Q algorithm, which is a variant of the slotted aloha protocol
Air interface	Pulse width modulation (PWM) for Class 0 and Class 1	Pulse interval encoding (PIE-ASK), Miller, FM0
Data rate	40/80 Kbits for Class 0 and 70/140 bits for Class 1	40 to 640 Kbits
Distance	Less than 10 meters	Less than 10 meters
Frequency range	850–930 MHz	860 to 960 MHz
Security password	8 and 24-bit passwords, respectively, for Class 1 and Class 0	32 bits
Data write verification	No	Yes
Write speed (for 96-bit electronic product code)	Three tags per second	Minimum five tags per second

Gen 2 protocol provides great versatility in properties of physical layers and linking layer procedures to suit different conditions (Zhang et al., 2009). Gen 2 standard specifies the communication between the reader and the tag regardless of the particular frequency range within which the system operates, thereby providing developers with basic blueprint detailing how the system (readers and tags) should communicate. It also provides a robust mechanism, the EPC number, for unique ID which can handle the need for the trillions of unique IDs required by healthcare sector. The Gen 2 frequency range is between 860 and 960 MHz and it covers all international frequency spectrums. Factors such as interoperability of Gen 2, short reading range and small tag size fulfil the requirement that the proposed passive RFID operates within this frequency range.

2.4.2 Active RFID

Given the availability of an internal power source, the sensors can be embedded into the active tag in order to extend the capability of the RFID system for tracking and monitoring applications. For instance, the vibration sensor is embedded into an active RFID system for asset monitoring (Hamid et al., 2013) as demonstrated in the architecture design of an active tag in Figure 2.5.




Figure 2.5 Architecture of an active RFID tag (Hamid et al., 2013)

In addition to that, the capability of the active tag to establish communication with the reader enables the RFID system to be actively deployed in real-time applications such as those that are event-driven and for periodic monitoring in the healthcare system. Table 2.6 displays the capability of an active RFID system in comparison to a passive RFID system (Cisco, 2008).

Table 2.6 Active RFID versus passive RFID (Cisco, 2008)

Characteristics	Passive RFID	Active RFID
Tag power source	Energy transferred from the reader via RF	Internal to tag
Tag battery	No	Yes
Availability of tag power	Only within field of reader	Continuous
Required signal strength from reader to tag	Very high	Very low
Communication Range	Short-range (3m or less)	Long-range, more than 100 m depends on types
Sensor capability	Read and transfer sensor data only when tag is power by reader without date/time stamp	Continuously monitor and record sensor input with date/time stamp
Data storage	Small (typically 128 bits)	Large (128kB or larger)